



Certificate No.L0583-110830

財團法人全國認證基金會
Taiwan Accreditation Foundation

Certificate of Accreditation

This is to certify that

Chroma Ate Inc.

Temperature Calibration Laboratory

68, Hwa - Ya 1st Rd., Hwa - Ya Technical Park, Taoyuan Hsien 333, Taiwan (R.O.C.)

is accredited in respect of laboratory

Accreditation Criteria : ISO/IEC 17025:2005
Accreditation Number : 0583
Originally Accredited : November 15, 1999
Effective Period : November 15, 2011 to November 14, 2014
Accredited Scope : Calibration Field, see described in the Appendix

Jay-San Chen
President, Taiwan Accreditation Foundation
Date : August 30, 2011



Certificate No.L0583-110830

財團法人全國認證基金會
Taiwan Accreditation Foundation

Accreditation Number : 0583

Laboratory Head : CHIANG, Ming-Yao

Temperature/Humidity

Calibration items	Calibration Method /Working Standard	Ranges	Smallest uncertainty
KE1002 platinum resistance thermometer (Null below)	E-04-22-025 /Resistance Thermometer Sensor ASLF250/T100	-20 °C	0.08 °C
		0 °C	0.09 °C
		50 °C	0.08 °C
		100 °C	0.11 °C
		150 °C	0.20 °C
		200 °C	0.21 °C
		250 °C	0.32 °C
		300 °C	0.35 °C

Remarks : Smallest uncertainty represents an expanded uncertainty using a coverage factor approximately 95 % level of confidence.

Approval Signatory	Scope
CHIANG, Ming-Yao	KE1002